

## **Certificate of Analysis**

3/6/2023

Product No.: RS4060-A

**Description:** Avantik Labeled 10% Neutral Buffered Formalin, 2.5 gallon cube

**Lot No.:** 32280

**Date of Manufacturing:** 3/2/2023

**Expiration Date:** 3/1/2025

| 10% Formalin Lot #: 3C001  |   |             |
|----------------------------|---|-------------|
| PROPERTY                   | TEST SPECIFICATION                          | RESULTS     |
| PH                         | 6.90-7.10 @ 25°C                            | 7.02@ 25°C  |
| Formaldehyde Concentration | 9.5-10.5% (v/v)                             | 10.31%(v/v) |
| Appearance                 | Clear, colorless, no observable particulate | Conforms    |

**Country of Origin:** United States of America **Latex Free:** This product contains no latex