



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

CERTIFICATE OF ANALYSIS

**RS4245 - Bluing Reagent Classic 4x1
Gal Avantik (MTO)**

Lot Number: 188543

Expiration Date: 2027-January-31

COMPONENTS

Source: SOL-6600/188695

Test	Specifications	Results
Color	Clear	Clear
pH @ above 20 C	≥ 7.80 and ≤ 8.20	7.83
Specific Gravity (g/cm3)	≥ 0.980 and ≤ 1.015	1.006

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.