



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

CERTIFICATE OF ANALYSIS

Lot Number: 174651

**RS4363-B - Optik Bluing Solution Type 1,
RTU, 1 Gallon (MTO)**

Expiration Date: 2025-May-31

COMPONENTS

Source: SOL-BU0398/174785

| Test | Specifications | Results |
|-------|-----------------------------|----------------|
| Color | Match Standard | Match Standard |
| pH | ≥ 7.90 and ≤ 8.10 | 7.98 |

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.