



## CERTIFICATE OF ANALYSIS

**RS4365-A - Optik Eosin Type 2, 500mL  
(MTO)**

Lot Number: 181555

Expiration Date: 2025-March-31

### COMPONENTS

Source: SOL-BU0418/181700

Test	Specifications	Results
pH	$\geq 5$ and $\leq 5$	5
Specific Gravity	$\geq 1$ and $\leq 1$	1

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.