



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

CERTIFICATE OF ANALYSIS

Lot Number: 186332

**RS4365-A - Optik Eosin Type 2, 500mL
(MTO)**

Expiration Date: 2025-May-31

COMPONENTS

Source: SOL-BU0418/186491

Test	Specifications	Results
pH	≥ 5 and ≤ 5	5
Specific Gravity	≥ 1 and ≤ 1	1

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.