



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

CERTIFICATE OF ANALYSIS

**RS4368-B - Optik Bluing Reagent Type 2,
1 Gallon (MTO)**

Lot Number: 179618

Expiration Date: 2026-September-30

COMPONENTS

Source: SOL-6600/182780

Test	Specifications	Results
Color	Clear	Clear
pH @ above 20 C	≥ 7.80 and ≤ 8.20	8.01
Specific Gravity (g/cm3)	≥ 0.980 and ≤ 1.015	1.014

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.