



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

CERTIFICATE OF ANALYSIS

Lot Number: 186388

**RS4421-500 - Light Green Counterstain -
500mL (MTO)**

Expiration Date: 2025-January-31

COMPONENTS

Source: SOL-STLGC/186543

Test	Specifications	Results
Color	Dark green	Dark green
pH @ 20-30 C	≥ 2.50 and ≤ 3.50	3.10
UV/VIS Absorbance (A)	≥ 0.10 and ≤ 0.25	0.15
UV/VIS Wavelength (nm)	≥ 630 and ≤ 635	632

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.