



19 Chapin Road - Building C, Pine Brook, NJ 07058  
800.783.9424 | avantik-us.com

## CERTIFICATE OF ANALYSIS

Lot Number: 191843

**RS4421-500 - Light Green Counterstain -  
500mL (MTO)**

Expiration Date: 2025-April-30

### COMPONENTS

Source: SOL-STLGC/191969

Test	Specifications	Results
Color	Dark green	Dark green
pH @ 20-30 C	$\geq 2.50$ and $\leq 3.50$	2.94
UV/VIS Absorbance (A)	$\geq 0.10$ and $\leq 0.25$	0.16
UV/VIS Wavelength (nm)	$\geq 630$ and $\leq 635$	631

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.