



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

CERTIFICATE OF ANALYSIS

Lot Number: 191764

**RS4435-500 - Periodic Acid, 0.5%, 500mL
(MTO)**

Expiration Date: 2025-June-30

COMPONENTS

Source: SOL-AHP03/191902

Test	Specifications	Results
Color	Clear	Clear
pH @ 20-30 C	≥ 1.75 and ≤ 2.75	1.86
Specific Gravity	≥ 0.95 and ≤ 1.01	1.00

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.