



CERTIFICATE OF ANALYSIS

**RS4546-125 - Alcian Blue - PAS Stain Kit
- 125mL (MTO)**

Lot Number: 182088

Expiration Date: 2024-September-30

COMPONENTS

Source: SOL-8000/182787

Test	Specifications	Results
Color	Clear	Clear
pH @ above 20 C	≥ 0.00 and ≤ 2.20	1.62
Specific Gravity (g/cm ³)	≥ 0.98 and ≤ 1.05	1.00
Functional Test	Pass/Fail	Pass

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.