



CERTIFICATE OF ANALYSIS

**GR2056-BL - Tissue Marking Dye
BLUE 2 oz Avantik (MTO)**

Lot Number: 172404

Expiration Date: 2025-April-30

COMPONENTS

Source: SOL-9007A/172540

Test	Specifications	Results
Color	Blue	Blue
Viscosity @ 20 - 30 C (cps)	≥ 380 and ≤ 580	414
pH	≥ 7.0 and ≤ 10.0	8.8

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.