



CERTIFICATE OF ANALYSIS

**11-6700-24 - Hi-Lite / Acid Rinse 4x1 Gal
Avantik (MTO)**

Lot Number: 152679

Expiration Date: 2025-June-30

COMPONENTS

Source: SOL-6700/153330

Test	Specifications	Results
Color	Clear	Clear
pH @ above 20 C	≥ 2.10 and ≤ 2.50	2.50
Specific Gravity (g/cm3)	≥ 0.980 and ≤ 1.015	1.000

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.