



CERTIFICATE OF ANALYSIS

11-5104-14 - Stain Harris Hematoxylin
Avantik 4x32 oz (MTO)

Lot Number: 151711

Expiration Date: 2023-December-31

COMPONENTS

Source: SOL-5104/151112

Test	Specifications	Results
Color	Red-Purple	Red-Purple
pH @ above 20 C	≥ 2.30 and ≤ 2.45	2.32
Specific Gravity (g/cm ³)	≥ 0.98 and ≤ 1.10	1.04
UV/VIS Absorbance (A)	≥ 0.95 and ≤ 1.18	0.98
UV/VIS Wavelength (nm)	≥ 555 and ≤ 560	557
Functional Test	Pass/Fail	N/A

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.