



CERTIFICATE OF ANALYSIS

**11-5103-14 - Stain Gill 3 Hematoxylin
Avantik 4x32 oz (MTO)**

Lot Number: 152247

Expiration Date: 2023-December-31

COMPONENTS

Source: SOL-5103/151144

Test	Specifications	Results
Color	Purple-Red	Purple-Red
pH @ above 20 C	≥ 2.05 and ≤ 2.35	2.34
Specific Gravity (g/cm ³)	≥ 0.98 and ≤ 1.10	1.04
UV/VIS Absorbance (A)	≥ 1.03 and ≤ 1.23	1.13
UV/VIS Wavelength (nm)	≥ 555 and ≤ 560	556
Functional Test	Pass/Fail	N/A

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.