



## CERTIFICATE OF ANALYSIS

**RS4028-B - Gill's Hematoxylin 1, 1 Gallon  
(MTO)**

Lot Number: 147180

Expiration Date: 2023-October-31

### COMPONENTS

Source: SOL-5101/147267

Test	Specifications	Results
Color	Purple	Purple
pH @ above 20 C	$\geq 2.35$ and $\leq 2.50$	2.43
Specific Gravity (g/cm <sup>3</sup> )	$\geq 0.98$ and $\leq 1.10$	1.05
UV/VIS Absorbance (A)	$\geq 0.48$ and $\leq 0.64$	0.52
UV/VIS Wavelength (nm)	$\geq 555$ and $\leq 560$	558
Functional Test	Pass/Fail	N/A

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.