



CERTIFICATE OF ANALYSIS

**RS4028-B - Gill's Hematoxylin 1, 1 Gallon
(MTO)**

Lot Number: 153526

Expiration Date: 2024-January-31

COMPONENTS

Source: SOL-5101/153700

Test	Specifications	Results
Color	Purple	Purple
pH @ above 20 C	≥ 2.35 and ≤ 2.50	2.48
Specific Gravity (g/cm ³)	≥ 0.98 and ≤ 1.10	1.04
UV/VIS Absorbance (A)	≥ 0.48 and ≤ 0.64	0.48
UV/VIS Wavelength (nm)	≥ 555 and ≤ 560	557
Functional Test	Pass/Fail	N/A

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.