



CERTIFICATE OF ANALYSIS

**RS4243 - 37% Formaldehyde Avantik 1
gal (MTO)**

Lot Number: 172365

Expiration Date: 2025-April-30

COMPONENTS

Source: 11-1400-25/172365

Test	Specifications	Results
Color	Clear	Clear
APHA (Pt-Co units)	≥ 0 and ≤ 10	5
Methanol %	≥ 10.0 and ≤ 15.0	12.4
Formaldehyde Concentration (%)	≥ 36 and ≤ 37	37

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.