



CERTIFICATE OF ANALYSIS

**RS4362-A - Optik Bluing Solution Type1,
Concentrate, 500mL, 4/cs**

Lot Number: 144999

Expiration Date: 2024-March-31

COMPONENTS

Source: SOL-BU0399/145032

Test	Specifications	Results
pH @ above 22C	≥ 7.90 and ≤ 8.20	8.07
Specific Gravity	Report	1.015

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.