



CERTIFICATE OF ANALYSIS

**RS4366-B - Optik Aqueous Clarifier Type
2, 1 Gallon**

Lot Number: 143384

Expiration Date: 2025-February-28

COMPONENTS

Source: SOL-6700/143448

Test	Specifications	Results
Color	Clear	Clear
pH @ above 20 C	≥ 2.10 and ≤ 2.50	2.46
Specific Gravity (g/cm3)	≥ 0.980 and ≤ 1.015	1.000

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.