



CERTIFICATE OF ANALYSIS

**RS4366-B - Optik Aqueous Clarifier Type
2, 1 Gallon**

Lot Number: 155282

Expiration Date: 2025-August-31

COMPONENTS

Source: SOL-6700/155453

Test	Specifications	Results
Color	Clear	Clear
pH @ above 20 C	≥ 2.10 and ≤ 2.50	2.44
Specific Gravity (g/cm3)	≥ 0.980 and ≤ 1.015	0.995

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.