



## CERTIFICATE OF ANALYSIS

RS4388-500 - Alcian Blue Stain, pH 2.5,  
500mL(MTO)

Lot Number: 155590

Expiration Date: 2024-January-31

### COMPONENTS

Source: SOL-5001/155000

Test	Specifications	Results
Color	Mid Blue	Mid Blue
Temperature (Celcius)	$\geq 20$	22
pH	$\geq 2.45$ and $\leq 2.55$	2.45
Specific Gravity (g/cm3)	$\geq 0.99$ and $\leq 1.02$	1.00
UV/VIS Absorbance (A)	$\geq 1.70$ and $\leq 1.90$	1.86
UV/VIS Wavelength (nm)	$\geq 620$ and $\leq 625$	625

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.