



## CERTIFICATE OF ANALYSIS

**RS4604-B - Rapid Decalcifier, 1 Gallon  
(MTO)**

Lot Number: 147697

Expiration Date: 2025-April-30

### COMPONENTS

Source: SOL-1212/148072

Test	Specifications	Results
Color	Clear	Clear
pH @ 21 - 30 C	$\geq 0.00$ and $\leq 1.00$	0.00
Specific Gravity (g/cm3)	$\geq 1.02$ and $\leq 1.05$	1.04

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.